



## Request for Proposal for Supply of Medical Oxygen Flow Meters

Reference	<b>RFP/CHAI-KE-EAPOA-010/25</b>
Posting Date	25th April 2025
Closing date	23rd May 2025

### Introduction

The Clinton Health Access Initiative (CHAI) is supporting the strengthening of medical oxygen ecosystems, including the improvement of medical oxygen availability and supply systems in Eastern Africa through support of Unitaid. The East Africa Program on Oxygen Access (EAPOA) aims to improve access to medical oxygen in low- and middle-income countries (LMICs) across East Africa. The initiative seeks to bolster the capacity of local manufacturers to meet the growing demand for medical oxygen.

CHAI intends to use part of the investment from unitaid to cover for Procurement of Medical oxygen flow meters.

CHAI invites financial proposals (RFP) from Manufacturers interested and capable of supplying medical oxygen flow meters into Kenya.

The purpose of this Request for Proposal is to:

- Provide information to potential Manufacturers on CHAI's needs and requirements.
- Solicit information from potential manufacturers on their product offerings, capacity, current lead times, pricing, and other information pertinent to procurement planning.
- Evaluate submissions to identify and shortlist qualified potential Manufacturers

### Submissions

Interested manufacturers must submit their RFP (electronically) to the email ID [chaikebids@clintonhealthaccess.org](mailto:chaikebids@clintonhealthaccess.org) quoting RFP/CHAI-KE-EAPOA-008/25 in the subject line no later than **23rd May 2025**.

For any clarification / queries please write to us at [chaikebids@clintonhealthaccess.org](mailto:chaikebids@clintonhealthaccess.org) with RFP reference number in the subject line by **9th May 2025**. CHAI may contact respondents for further discussions regarding submissions and subsequent procurement.

Should your submission exceed 2Mb and cannot be sent by email, send an email to request an upload link with the subject 'Requesting upload link' to [chaikebids@clintonhealthaccess.org](mailto:chaikebids@clintonhealthaccess.org).

**Submissions received after the deadline will not be considered.**

#### **Disclaimer**

Distribution of this document does not mean there is any commitment on the part of CHAI to engage any entity. CHAI will not reimburse or otherwise bear any costs associated with this RFP. No fee is charged for the submission of this RFP. All material received in the response to this RFP shall become the property of CHAI and will not be returned to the respondent. CHAI reserves the right to use any information presented in any proposal.

#### **CHAI code of conduct for suppliers**

The CHAI code of conduct for suppliers is a minimum set of requirements that suppliers must meet to qualify to do business with CHAI and remain in good standing with the organization. Please read the document via this link: [Supplier Code of Conduct](#) and apply for this solicitation only if your organization can comply with these requirements.

## 1. Medical Oxygen Flow meters

Item	Specifications	QTY	Unit Rate \$	Total \$
1	<p><b>Design Parameters:</b>  Medical oxygen metering device with a Flow selector intended for the administration of the medical gases (Oxygen) in the treatment and care of patients.  Available for paediatric and neonatal version.</p> <p>Twin/Dual flow Configuration.</p> <p><b>Flow Versions</b>  1. Micro Flow – 0-1 lpm -Single flow  2. Standard flow – 0-15 lpm - Twin /Duo Flow  3. Autoclavable Humidifier Bottles</p> <p><b>Regulatory:</b>  Complies with Medical Devices Directive 93/42/EEC  Complies with Standard EN 1789  • Comply with EN ISO 15002.</p> <p><b>Warranty:</b>  Provide two years warranty</p> <p><b>Lead Times</b>  <i>Manufacturers to provide a delivery Schedule</i></p> <p><i>Proof of Supply</i>  Provide proof of supply to at least three clients with flow meters in Eastern Africa.</p>	200 100 600		

## 2. Summary Price Proposal

Item	Description	Units	QTY	Unit \$	Total \$
1.	1. Micro Flow – 0-1 lpm -Single flow	No.	200		
	2. Standard flow – 0-15 lpm- Duo/Twin Flow	No.	100		
	3. Autoclavable Humidifier Bottles	No.	600		
	Sub Total				
	Less Discount				
	<b>Total</b>				